


**300 Blk E. College Ave. E/B**



**RADAR LAB OF MARYLAND**

318 St. Patrick Road

Baltimore, MD 21206 ■ (410) 866-2974

radarlab@outlook.com ■ <http://radarlabofmaryland.com>

**Certificate of Calibration**

Expiration: 07/08/2023

**Manufacturer: Smartmicro**

**Model: UMRR0A0303**

**Serial Number: 0x000230CE**

I hereby certify under penalty of perjury the following Speed Measuring Radar Device has been checked for accuracy and correctness of operation under my supervision. This Speed Measuring Radar Device is certified accurately within +/-1 mph in stationary mode using equipment with accuracy traceable to National Institute of Standards and Technology.

The Transmitter of this device has been tested and found to be within specified range for Radar Devices as established by the Federal Communications Commission.

**Microwave Frequency in measured: 24.15 GHz**

07/08/2022

Date

Technician Signature

At Thompson  
Technician Signature/NARTE Certification No. EI-03190


IN WITNESS HEREOF, I have set my hand and seal this 8<sup>th</sup> day of July, 2022. My commission expires \_\_\_\_\_

DELORES M PAYNE  
Notary Public-Maryland  
Howard County  
My Commission Expires  
August 01, 2024

Delores M. Payne  
Notary Public



**400 Blk E. College Ave. W/B**



**RADAR LAB OF MARYLAND**

318 St. Patrick Road

Baltimore, MD 21206 ■ (410) 866-2974

radarlab@outlook.com ■ <http://radarlabofmaryland.com>

**Certificate of Calibration**

Expiration: 07/08/2023

**Manufacturer: Smartmicro**

**Model: UMRR0A0303**

**Serial Number: 0x000230CD**

I hereby certify under penalty of perjury the following Speed Measuring Radar Device has been checked for accuracy and correctness of operation under my supervision. This Speed Measuring Radar Device is certified accurately within +/-1 mph in stationary mode using equipment with accuracy traceable to National Institute of Standards and Technology.

The Transmitter of this device has been tested and found to be within specified range for Radar Devices as established by the Federal Communications Commission.

**Microwave Frequency in measured: 24.15 GHz**

07/08/2022

Date

Technician Signature

Thompson  
Technician Signature/NARTE Certification No. EI-03190


IN WITNESS HEREOF, I have set my hand and seal this 8<sup>th</sup> day of July, 2022. My commission expires \_\_\_\_\_.

DELORES M. PAYNE  
Notary Public-Maryland  
Howard County  
My Commission Expires  
August 01, 2024

DeLores M. Payne  
Notary Public



**700 Blk Riverside Dr. N/B**



**RADAR LAB OF MARYLAND**

318 St. Patrick Road

Baltimore, MD 21206 ■ (410) 866-2974

radarlab@outlook.com ■ <http://radarlabofmaryland.com>

**Certificate of Calibration**

Expiration: 07/08/2023

**Manufacturer: Smartmicro**

**Model: UMRR0A0303**

**Serial Number: 0x00022A7F**

I hereby certify under penalty of perjury the following Speed Measuring Radar Device has been checked for accuracy and correctness of operation under my supervision. This Speed Measuring Radar Device is certified accurately within +/-1 mph in stationary mode using equipment with accuracy traceable to National Institute of Standards and Technology.

The Transmitter of this device has been tested and found to be within specified range for Radar Devices as established by the Federal Communications Commission.

**Microwave Frequency in measured: 24.15 GHz**

07/08/2022

Date

Technician Signature

  
Technician Signature/NARTE Certification No. EI-03190

IN WITNESS HEREOF, I have set my hand and seal this 8<sup>th</sup> day of July, 2022. My commission expires

DELORES M PAYNE  
Notary Public-Maryland  
Howard County  
My Commission Expires  
August 01, 2024

  
Notary Public



**600 Blk Business Rt. 50 E/B**



**RADAR LAB OF MARYLAND**

318 St. Patrick Road  
Baltimore, MD 21206 ■ (410) 866-2974  
radarlab@outlook.com

**Certificate of Calibration**

Expiration: 04/28/2024

Manufacturer: Smartmicro

Model: UMRR-0A0303

Serial Number: 0x000230D3

I hereby certify under penalty of perjury the following Speed Measuring Radar Device has been checked for accuracy and correctness of operation under my supervision. This Speed Measuring Radar Device is certified accurately within +/-1 mph in stationary mode using equipment with accuracy traceable to National Institute of Standards and Technology.

The Transmitter of this device has been tested and found to be within specified range for Radar Devices as established by the Federal Communications Commission.

Frequency: 24.15 GHz

Out Put Power: 20dbm

04/28/2023

Date

Technician Signature

Thompson  
Technician Signature/NARTE Certification No. EI-03190


IN WITNESS HEREOF, I have set my hand and seal this 28<sup>th</sup> day of April, 2023. My commission expires

DELORES M PAYNE  
Notary Public-Maryland  
Howard County  
My Commission Expires  
August 01, 2024

Delores M Payne  
Notary Public



**600 Blk Business Rt. 50 W/B**



**RADAR LAB OF MARYLAND**

318 St. Patrick Road

Baltimore, MD 21206 ■ (410) 866-2974

radarlab@outlook.com ■ <http://radarlabofmaryland.com>

**Certificate of Calibration**

Expiration: 07/08/2023

**Manufacturer: Smartmicro**

**Model: UMRR0A0303**

**Serial Number: 0x00022C4B**

I hereby certify under penalty of perjury the following Speed Measuring Radar Device has been checked for accuracy and correctness of operation under my supervision. This Speed Measuring Radar Device is certified accurately within +/-1 mph in stationary mode using equipment with accuracy traceable to National Institute of Standards and Technology.

The Transmitter of this device has been tested and found to be within specified range for Radar Devices as established by the Federal Communications Commission.

**Microwave Frequency in measured: 24.15 GHz**

07/08/2022

Date

Technician Signature

  
Technician Signature/NARTE Certification No. EI-03190

IN WITNESS HEREOF, I have set my hand and seal this 8<sup>th</sup> day of July, 2022. My commission expires \_\_\_\_\_.

DELORES M PAYNE  
Notary Public-Maryland  
Howard County  
My Commission Expires  
August 01, 2024

  
Notary Public



**700 Blk E. Main St. N/B**



**RADAR LAB OF MARYLAND**

318 St. Patrick Road  
Baltimore, MD 21206 ■ (410) 866-2974  
radarlab@outlook.com

**Certificate of Calibration**

Expiration: 04/28/2024

**Manufacturer: Smartmicro**

**Model: UMRR-0A0303**

**Serial Number: 0x000230BD**

I hereby certify under penalty of perjury the following Speed Measuring Radar Device has been checked for accuracy and correctness of operation under my supervision. This Speed Measuring Radar Device is certified accurately within +/-1 mph in stationary mode using equipment with accuracy traceable to National Institute of Standards and Technology.

The Transmitter of this device has been tested and found to be within specified range for Radar Devices as established by the Federal Communications Commission.

**Frequency: 24.15 GHz**

**Out Put Power: 20dbm**

04/28/2023

Date

Technician Signature

Technician Signature/NARTE Certification No. EI-03190

IN WITNESS HEREOF, I have set my hand and seal this 28<sup>th</sup> day of April, 2023. My commission expires \_\_\_\_\_.

DELORES M PAYNE  
Notary Public-Maryland  
Howard County  
My Commission Expires  
August 01, 2024

Notary Public